

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/03/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:28
INTERNAL STANDARD	VERIFIED	05:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:29
BLANK TEST	0.00	05:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:30
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:31
BLANK TEST	0.00	05:32

