

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/04/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

0.00	05:40	BLANK TEST
0.00	05:40	INTERNAL STANDARD
0.09	05:41	REFERENCE STANDARD
0.00	05:42	BLANK TEST
0.09	05:42	REFERENCE STANDARD
34.01c	05:43	REFERENCE STANDARD TEMP
0.00	05:43	BLANK TEST
0.09	05:43	REFERENCE STANDARD
34.01c	05:43	REFERENCE STANDARD TEMP
0.09	05:43	REFERENCE STANDARD
0.00	05:44	BLANK TEST

