

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/16/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:45
INTERNAL STANDARD	VERIFIED	01:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49

