

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/23/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:28
INTERNAL STANDARD	VERIFIED	17:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:29
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:30
BLANK TEST	0.00	17:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:31
BLANK TEST	0.00	17:32

