

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/06/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:46
INTERNAL STANDARD	VERIFIED	17:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:49
BLANK TEST	0.00	17:49

