

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 02/12/2025

OPERATOR NAME:  
ANTHONY J TRAVIGLIA  
Permit #: 53024  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41

