

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/16/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:01
INTERNAL STANDARD	VERIFIED	19:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:01
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:02
BLANK TEST	0.00	19:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:04
BLANK TEST	0.00	19:04

