

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 01/01/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:19
INTERNAL STANDARD	VERIFIED	21:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:20
BLANK TEST	0.00	21:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:22
BLANK TEST	0.00	21:23

