

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:55
INTERNAL STANDARD	VERIFIED	05:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:57
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:58
BLANK TEST	0.00	05:59

