

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/02/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50815
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:01
INTERNAL STANDARD	VERIFIED	18:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:01
BLANK TEST	0.00	18:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:02
BLANK TEST	0.00	18:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:03
BLANK TEST	0.00	18:04

