

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 03/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:40
INTERNAL STANDARD	VERIFIED	05:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44

