

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

cfo

Date: 02/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:49
INTERNAL STANDARD	VERIFIED	05:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:51
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:53

