

ONEIDA CO.
SHERIFF'S DEPT
ATAMASTER dmt: 123906

fd

ate: 06/15/2026

PERATOR NAME:
AILEY J DIXON
ermit #: 50814
perator Agency/Dept: ONEIDA CO. SO
eference Standard #: 26120

ominal = 0.10 g/210L

BLANK TEST	0.00	17:55
INTERNAL STANDARD	VERIFIED	17:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59

