

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/17/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25120

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:26
INTERNAL STANDARD	VERIFIED	04:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:27
BLANK TEST	0.00	04:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:28
BLANK TEST	0.00	04:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:29
BLANK TEST	0.00	04:30

