

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/04/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:30
INTERNAL STANDARD	VERIFIED	17:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:32
BLANK TEST	0.00	17:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:33
BLANK TEST	0.00	17:34

