

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/16/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:27
INTERNAL STANDARD	VERIFIED	17:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:28
BLANK TEST	0.00	17:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:29
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:30
BLANK TEST	0.00	17:31

