

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/07/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:21
INTERNAL STANDARD	VERIFIED	01:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25

