

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/31/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:45
INTERNAL STANDARD	VERIFIED	13:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:46
BLANK TEST	0.00	13:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:47
BLANK TEST	0.00	13:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:48
BLANK TEST	0.00	13:49

