

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/02/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:23
INTERNAL STANDARD	VERIFIED	17:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:23
BLANK TEST	0.00	17:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:24
BLANK TEST	0.00	17:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:26
BLANK TEST	0.00	17:26

