

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/16/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: 0320000
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:51
INTERNAL STANDARD	VERIFIED	16:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:52
BLANK TEST	0.00	16:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:53
BLANK TEST	0.00	16:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:54
BLANK TEST	0.00	16:55

