

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/24/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:22
INTERNAL STANDARD	VERIFIED	17:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:22
BLANK TEST	0.00	17:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:24
BLANK TEST	0.00	17:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:25
BLANK TEST	0.00	17:26

