

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/05/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:31
INTERNAL STANDARD	VERIFIED	17:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:32
BLANK TEST	0.00	17:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:34
BLANK TEST	0.00	17:34

