

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 03/04/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48

