

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

*Cfb*

Date: 04/08/2026

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:59
INTERNAL STANDARD	VERIFIED	01:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02

