

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

efo

Date: 04/15/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:29
INTERNAL STANDARD	VERIFIED	17:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:30
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:32
BLANK TEST	0.00	17:33

