

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/11/2026

cf

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:48
INTERNAL STANDARD	VERIFIED	01:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52

