

SUPERVISOR TEST




ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/08/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:05
INTERNAL STANDARD	VERIFIED	04:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:06
BLANK TEST	0.00	04:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:08
BLANK TEST	0.00	04:09



CF0

[Signature] 66

7/8/26