

SUPERVISOR TEST




ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/12/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:35
INTERNAL STANDARD	VERIFIED	22:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39



CFO

BW 66