

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

*cfo*

Date: 02/18/2026

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:35
INTERNAL STANDARD	VERIFIED	17:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:35
BLANK TEST	0.00	17:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:38
BLANK TEST	0.00	17:39

