

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 02/25/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:01
INTERNAL STANDARD	VERIFIED	21:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:02
BLANK TEST	0.00	21:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:03
BLANK TEST	0.00	21:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:04
BLANK TEST	0.00	21:05

CFO

