

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/29/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:47
INTERNAL STANDARD	VERIFIED	18:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:47
BLANK TEST	0.00	18:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:51

