

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/20/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:09
INTERNAL STANDARD	VERIFIED	05:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:10
BLANK TEST	0.00	05:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:11
BLANK TEST	0.00	05:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:12
BLANK TEST	0.00	05:13

