

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/25/2026

CPD

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:18
INTERNAL STANDARD	VERIFIED	02:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:21

