

SOLUTION CHANGE

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606  
-----

Date: 09/01/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:47
INTERNAL STANDARD	VERIFIED	18:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:48
BLANK TEST	0.00	18:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:51

