

SOLUTION CHANGE

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/01/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:11
INTERNAL STANDARD	VERIFIED	02:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:11
BLANK TEST	0.00	02:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:14

