

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/25/2021

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:12
INTERNAL STANDARD	VERIFIED	22:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:15
BLANK TEST	0.00	22:16

