

**SUPERVISOR TEST**

-----  
 ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606  
 -----

Date: 09/22/2021

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:18
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:21
BLANK TEST	0.00	18:22

