

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 09/29/2021

OPERATOR NAME:  
STEVEN P EILRERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:10
INTERNAL STANDARD	VERIFIED	18:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:14

