

LEB

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/07/2021

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:10
INTERNAL STANDARD	VERIFIED	04:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:12
BLANK TEST	0.00	04:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:14

