

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/27/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:53
INTERNAL STANDARD	VERIFIED	17:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:55
BLANK TEST	0.00	17:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:56
BLANK TEST	0.00	17:57

