

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/03/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:59
INTERNAL STANDARD	VERIFIED	05:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:01
BLANK TEST	0.00	06:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:02
BLANK TEST	0.00	06:03

