

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/24/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:10
INTERNAL STANDARD	VERIFIED	23:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:11
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:12
BLANK TEST	0.00	23:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:13
BLANK TEST	0.00	23:14

