

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/08/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:36
INTERNAL STANDARD	VERIFIED	18:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:36
BLANK TEST	0.00	18:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:38
BLANK TEST	0.00	18:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:39
BLANK TEST	0.00	18:40

