

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/15/2021

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:05
BLANK TEST	0.00	22:06

