

SOLUTION CHANGE

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/01/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:24
INTERNAL STANDARD	VERIFIED	18:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:26
BLANK TEST	0.00	18:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:27
BLANK TEST	0.00	18:28

