

SOLUTION CHANGE

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

LEB

Date: 12/01/2022

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:36
INTERNAL STANDARD	VERIFIED	03:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40

