

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 07/20/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:42
INTERNAL STANDARD	VERIFIED	04:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:44
BLANK TEST	0.00	04:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:46

