

SUPERVISOR TEST

LER

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 08/24/2022

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:53
INTERNAL STANDARD	VERIFIED	17:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:54
BLANK TEST	0.00	17:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:56
BLANK TEST	0.00	17:56

