

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 10/19/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:59
INTERNAL STANDARD	VERIFIED	22:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:01
BLANK TEST	0.00	23:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:02
BLANK TEST	0.00	23:03

