

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 11/23/2022

OPERATOR NAME:  
NOAH A DANQUER  
Permit #: 50812  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:00
INTERNAL STANDARD	VERIFIED	06:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:01
BLANK TEST	0.00	06:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:02
BLANK TEST	0.00	06:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:03
BLANK TEST	0.00	06:04

LEB

